



A SIMPLE SPECTROPHOTOMETRIC METHOD FOR THE DETERMINATION OF COBALT (II) USING 1-NITROSO-2-NAPHTHOL IN ANIONIC AQUEOUS SOLUTION OF SODIUM DODECYL SULPHATE

GUL AFSHAN SOOMRO^a and GHULAM ABBAS SHAR^{*,a,b}

^aDepartment of Chemistry, Shah Abdul Latif University, KHAIRPUR – 66022, Sindh, PAKISTAN

^bDepartment of Chemistry, Virginia Commonwealth University, Richmond,
VIRGINIA – 23284-2006, USA

ABSTRACT

A simple rapid new spectrophotometric method has been developed for the determination of cobalt at trace level using 1-nitroso-2-naphthol (NNPh) in presence of anionic aqueous micellar solution surfactant 1.0% sodium dodecyl sulphate (SDS). NNPh reacts with cobalt (II) to form green coloured tris (1-nitroso-2-naphtholato) cobalt complex. The importance of this method is the use of micellar system which replaces the previous solvent extraction steps while enhancing the sensitivity, selectivity and the molar absorptivity which reduces the cost and toxicity. The average molar absorption coefficient and Sandell's sensitivity was found to be $2.05 \times 10^4 \text{ L mol}^{-1}\text{cm}^{-1}$ and 3.49 ng cm^{-2} at λ_{max} 436.2 nm. Linear calibration graph was obtained over the concentration range $0.12\text{-}4.0 \mu\text{g mL}^{-1}$, the stoichiometric composition of the chelate was 1:3 (Co: [NNPh]₃). The developed method was successfully applied for the determination of cobalt from alloy and pharmaceutical samples.

Key words: Spectrophotometric, Cobalt, NNPh, Anionic micellar, SDS.

INTRODUCTION

Cobalt play essential role in biological systems¹. Cobalt is the major constituent of vitamin B₁₂ or cyanocobalamine. Cobalt deficiency causes thalassemia, sickle cell anemia, hematological disorders, and polycythemia disease². Excess dose of cobalt is toxic and causes vomiting, nausea, pulmonary diseases and skin diseases³. Special cobalt- alloys are used for prosthetic parts such as knee and hip replacements⁴. Cobalt ion in small concentration is essential for enzymatic functions in all mammals. Cobalt is very toxic, when it is given to plants and moderately toxic to mammals in high dose. Now days, it has been

* Author for correspondence; E-mail: gabbas.shar@salu.edu.pk

great concern about the role of cobalt in biochemical and environmental systems. Determination of cobalt concentration and its effects on system is of great importance and scope. In spectrophotometric analysis of metal ions, the metal is complexed with chelating agent and solubilized in solvent extraction is easy method for studying the metal ions because of its simplicity and rapidity⁵. Several spectrophotometric methods have been developed to replace the old solvent extraction method by using the surfactants⁶. Micellar methods due to solubility of several metal complexes have shown improvement in the analytical merits of techniques for the determination of metal ions. Micellar media is introduced to enhance the molar absorptivities, sensitivities and replaces the previous steps of toxic solvent extraction⁷⁻¹⁵. Many spectrophotometric methods for the determination of cobalt using various chelating reagents are available. Some of the recently developed spectrophotometric methods for the determination¹⁶⁻²¹ are less selective and less sensitive. We have developed new simple, sensitive and selective spectrophotometric method for the determination of cobalt (II) in various materials by using 1-NNPh as chromogenic agent in anionic micellar aqueous solution of 1.0% SDS.

EXPERIMENTAL

Apparatus

A Cecil CE 9500 UV-Vis scanning spectrophotometer equipped with a 10-mm path length quartz cell was utilized. Atomic absorption spectrophotometer (Analyst-100 Perkin Elmer) with multi-elemental hollow cathode lamps was used. I.R spectrophotometer (260 Hitachi) was utilized for study of metal chelates. Besides, pH/conductivity meter (Sension 156 HACH) with Sension Gel-filled pH electrodes was used.

Preparation of reagents

NNPh (Merck) was made by dissolving 75 mg in 50 mL and was dissolved in minimum volume of 1.0% SDS, this was stock solution of 15000 ppm ($1500 \mu\text{g mL}^{-1}$). SDS 1.0% was prepared by taking 1.0 g in volumetric flask of 100 mL. Cobalt (II) nitrate was made prepared by dissolving 4.93 g in ($1000 \mu\text{g mL}^{-1}$) 1L. Copper (II) nitrate stock solution was prepared by taking 3.85 g in ($1000 \mu\text{g mL}^{-1}$) 1L. Nickel (II) nitrate was made by dissolving 4.94 g in ($1000 \mu\text{g mL}^{-1}$) 1L and iron (II) chloride was prepared by taking 4.84 g in 1 L ($1000 \mu\text{g mL}^{-1}$). Other metal ions solutions were made by dissolving their nitrate and chloride salts and insoluble compounds were made dissolved in acids or according to special dissolution method²². Buffer solutions of pH 1-10 were prepared by using appropriate methods²³, with appropriate volumes of the mixtures of 0.2 M KCl-0.2 M HCl for (pH 1-4.0), 0.2 M CH_3COOH -0.2 M CH_3COONa for (pH 5-6.0), 0.1 M KH_2PO_4 -0.1 M NaOH for (pH 6.5-8.0) and 0.025 M sodium borate and 0.1 M HCl for (pH 9-10.0).

Table 1: Analytical parameters of Co(II)- [NNPh]₃ in 1.0% SDS

Parameters	Co(II)
Wavelength, λ_{\max} nm	436.2
pH	5
SDS	1.0 %
NNPh (μgmL^{-1})	185
Reagent (M:R)	1:3
Linear range/ μgmL^{-1}	0.12 -4
Molar absorption coefficient/ $\text{mol}^{-1}\text{cm}^{-1}$	2.05×10^4
Sandell's sensitivity/ ngcm^{-2}	3.49
Detection limit/ μgmL^{-1}	0.12
Correlation coefficient (R^2)	0.999
Reproducibility (% RSD)*	0.91

Table 2: Effect of foreign ions on the Co(II)-[NNPh]₃

Ion/salt	Co(II)
Na ₂ tartarate, KSCN, KClO ₃	800
Mg(II)	600
Na ₂ C ₂ O ₄	200
Na ₂ citrate	500
Mn(II)	200
Zn(II)	50
Cu(II)	5
KCN	50
Al(III)	10
Ni(II)	5
Cr(IV)	8
Fe(III)	5
Pb(II)	3
Cd(II)	2

Determination of cobalt ions in alloy sample

Alloy sample of cobalt ions 0.1-0.5 g of reference material JSS 607-6 was digested with concentrated HCl (15 mL) and concentrated HNO₃ (5.0 mL) in beakers. The sample solution was heated to reduce the volume to 5 mL on a hot plate. Then mixed 10 mL concentrated HCl, solutions in the sample solutions by filtering and diluting to a final volume of 25 mL. Alloy sample of reference material having the composition as JSS 607-6 (Co 4.72, W 16.96, V 0.82)²⁴ data values are shown in Table 3.

Table 3: Determination of Co(II) in reference material

Certified (%) composition	Metal ion	Metal (µg) present	Metal (µg) found	RSD %	Relative error %	Recovery %
JSS 607-6	Co(II)	14.01	14.03	0.80	0.14	100.14

Pharmaceutical samples

Neurobion injection (Merck) (vitamin B₁₂) sample 1.0 mL ampoule injection was transferred into 50 mL flask adding 10 mL nitric acid and sulphuric acid ratio (10+1), sample was heated to dryness²⁵. The sample residue was leached with dilute H₂SO₄ and diluted to the solution till it reaches final colorless residue²⁶. Sample solution was neutralized with sodium hydroxide and was made working solution in 50 mL volumetric flask. However, it was transferred in volumetric flask, and added adequate volume of NNPh 185 ppm, 1.0% SDS and pH 5 buffer solutions for the complex formation. The results are given in Table 4.

Table 4: Determination of Co(II) in pharmaceutical samples

Sample	Analyte ion	Proposed method µgmL ⁻¹	RSD %	AAS method µgmL ⁻¹	RSD %	Recovery %
Neurobion (Inj) 21.74 µg mL ⁻¹	Co(II)	21.70	0.5	21.71	0.8	99.81

Determination of Co (II) in tap water samples

Sample solution of tap water was collected from Khairpur city. Then samples were filtered through 0.45 µm filter paper and then were acidified with 1 mL concentrated HNO₃ to avoid precipitation. Cobalt metal was skipped in tap water in the volumetric flask added appropriate volumes of NNPh 185 ppm, 2.0 mL buffer solution of pH 5 and 2 mL 1.0% SDS the complex absorbance was measured as given in Table 5.

Table 5: Percent recovery of known amount of Co(II) added to tap water

Metal ions	Amount added ($\mu\text{g mL}^{-1}$)	Amount found ($\mu\text{g mL}^{-1}$)	Recovery (%)
Co(II)	1.0	0.99	99

Table 6: Comparison of cobalt determination methods using NNPh

Metal	Method	Reference
Co	In Triton X-100, ϵ $3.18 \times 10^4 \text{ mol}^{-1} \text{ cm}^{-1}$, Sandell's sensitivity 2.05 ng mL^{-1} , linear range $0.20\text{-}3.0 \text{ } \mu\text{g mL}^{-1}$, and D.L 1.68 ng mL^{-1}	28
Co(II)	in Tween 80, D.L. $0.016 \text{ } \mu\text{g mL}^{-1}$	35
Co(II)	In Triton X-100, ϵ $2.05 \times 10^4 \text{ mol}^{-1} \text{ cm}^{-1}$, λ_{max} 425 nm D.L., $0.02 \text{ } \mu\text{g mL}^{-1}$	27
Co(II)	In 1.0% SDS, ϵ $2.05 \times 10^4 \text{ mol}^{-1} \text{ cm}^{-1}$, 436.2 nm , D.L. $3.49 \text{ } \mu\text{g mL}^{-1}$	Present method

RESULTS AND DISCUSSION

NNPh is an excellent colour forming chelating agent reacts with cobalt forms tris (1-nitroso-2-naphtholato) cobalt complex structure is shown in Fig. 1. UV-vis spectra of NNPh transitions ($n \rightarrow \pi^*$) from ligand-to ligand charge transfer (LLCT)²⁷⁻²⁹ of N-O molecule and C-O-H group at λ_{max} 377.0 nm is shown in Fig. 2. UV-vis spectra (1-nitroso-2-naphtholato) cobalt complex of green colour is formed at λ_{max} 436.2 nm charge transfer transitions ($n \rightarrow \pi^*$), from ligand-to metal (LMCT) ($p\pi\text{-}d\pi$) is shown in Fig. 3. Cobalt (II)-[NNPh]₃ complexes showing increasing concentration of cobalt ions in 1.0% SDS at (λ_{max} 436.2 nm) are shown in Fig. 4.

The stoichiometry of the complex was found 1:3 (Metal : Ligand) by Job's method³⁰ of continuous variation method of metal and chelate ratio³¹ forming tris[1-nitroso-2-naphtholato]cobalt shown in Fig. 5. NNPh concentration was varied at fixed concentration of 1 m.mole Co(II) ions and NNPh molar ratios 5.0 to 80.0 m.mole , where $20\text{-}50 \text{ mM}$ NNPh was optimized concentration found for the complex formation is shown in Fig. 4. For surfactant optimization, $1 \text{ } \mu\text{g mL}^{-1}$ Co(II)-chelate, 2 mL of 1.0% SDS showed constant maximum absorbance in 10 mL volume, this 1.0% SDS was optimized for the whole procedure, which value is greater than the cmc value $8.3 \times 10^{-3} \text{ M}$ than reported³². Co-[NNPh]₃ complex showed maximum constant absorbance at pH 5.0 , which was selected for subsequent studies as shown in Fig. 7.




Fig. 1: Structure of tris(1-nitroso-2-naphtholato) cobalt




Fig. 2: UV-vis spectrum of NNPh reagent at (λ_{\max} 377.0 nm) in 1.0% SDS




Fig. 3: UV-vis spectrum of Co(II)-[NNPh]₃ complex at (λ_{\max} 436.2 nm) in 1.0% SDS




Fig. 4: UV-vis spectra of Co(II)-[NNPh]₃ with increasing concentration at (λ_{max} 436.2 nm)




Fig. 5: Composition of Co(II)-[NNPh]₃ by mole ratio method (Job's plot of metal : ligand ratio)

Calibration graph showed linear concentration over the range $0.12\text{--}4.0\ \mu\text{g mL}^{-1}$, the correlation coefficients with experimental data was 0.999 as shown in Fig. 5. Molar absorptivity of the complex was found $(2.05 \times 10^4)\ \text{L mol}^{-1}\ \text{cm}^{-1}$ is given in Table 1 and showed significant improvement in the molar absorptivities^{33,34}. Detection limit and Sandell's sensitivity was found $0.12\ \mu\text{g mL}^{-1}$ and $3.49\ \text{ng cm}^{-2}$ as given in Table 1. The metal-complex showed absorbance maximum constant up to 5 min. At ordinary temperature, metal complex showed stable absorbance measurement, till 24 h.




Fig. 6: Effect of amounts of NNPh on the absorbance of Co(II)




Fig. 7: Effect of pH on the absorbance of Co (II)

Effect of foreign ions in the determination of cobalt

Above the amount $800 \mu\text{g mL}^{-1}$ KSCN, sodium tartarate and KClO_3 showed interference in the Co(II)-chelate formation. Cadmium(II), copper(II), lead(II), iron(II) and nickel(II) showed interference at low trace level concentration. Ascorbic acid and EDTA removed the interference of other metal ions and were used as masking agents. Fe(II) showed no interference at pH 5, because it interferes at optimized pH 1 by forming metal complex. Cu(II) solutions interference was removed by using buffer solution of NH_3 as masking agent. Cu(II) interference was completely removed by using their specific masking agents results are shown in Table 2.




Fig. 8: Calibration graph for Co(II)-[NNPh]₃ complex in 1.0% SDS

Validation of method to the determination of cobalt (II) ions

The method developed was validated to the determination of cobalt ions by present spectrophotometric method from certified reference material JSS 607-6 and in Neurobion injection and also standard addition method of % age recovery test was applied for the verification of the accuracy and the precision of the method. The average Co ions found (n 7) in the reference material with a precision as relative standard deviation RSD was upto 0.80 % is given in Table 3-5.

CONCLUSION

The results obtained in the determination of cobalt (II) were comparable with the recently reported methods, while the present method showed significant improvement in Beer's law linear concentration range than reported³³⁻³⁶. Sensitivity and the selectivity of the present method is higher than the reported expensive and toxic solvent extraction methods. The present method is simple, rapid, sensitive, selective, reproducible and non-extractive than existing methods of spectrophotometric analysis as given in Table 6.

ACKNOWLEDGEMENT

Research supported by the Higher Education Commission, Islamabad, Pakistan under award number 20-653/R & D. G. A. Shar acknowledges Higher Education Commission, Islamabad, Pakistan for postdoctoral fellowship at Virginia Commonwealth University, USA.

REFERENCES

1. A. Wahi and L. R. Kakkar, *Analytical. Sci.*, **10**, 509 (1994).
2. M. A. Jakupec, M. Galanski and B. K. Keppler, *Rev. Physiol. Biochem. Pharmacol.*, **146**, 1 (2003).
3. Y. P. Ho, S. C. F. AuYeung and K. K. W. To, *Med. Resource Rev.*, **23**, 633 (2003).
4. R. Michel, M. Nolte, M. Reich and F. Loer, *Archives of Orthopaedic and Trauma Surgery*, **110**, 2, 61 (1991).
5. M. P. S. Andres, M. L. Marina and S. Vera, *Talanta*, **41**, 179 (1994).
6. G. A. Olah, *Chem. Eng. News*, **54(6)**, 6 (1976).
7. G. Gumus, H. Filik, B. Demirata and Adnan, in, 4th AACD, Congress Proceeding Book, Menderes University, Kusadası-Aydin, Turkey (2004) p. 125.
8. N. K. Agnihotri, V. K. Singh, S. Ratnani, S. K. Shukla and G. K. Parashar, *Analytical Letters*, **38**, 12, 1999 (2005).
9. R. K. Agarwal, D. Sharma, L. Singh and H. Agarwal, *Bioinorg. Chem. Appl.*, doi. 10.1155/BCA/2006/29234 (2006).
10. M. Nako, C. Yamazaki, H. Tominaga, T. Yamaguchi and Y. Fujita, *Anal. Sci.*, **22(2)**, 313 (2006).
11. H. Khan, M. J. Ahmad and M. I. Bhangar, *Spectroscopy*, **20(5-6)**, 285 (2006).
12. K. Varinder, M. A. Kumar and V. Neelam, *Annali di Chimica*, **97(3-4)**, 237 (2007).
13. G. A. Shar and M. I. Bhangar, *J. Chem. Soc. Pakistan*, **25(1)**, 28 (2003).
14. H. Khan, M. J. Ahmed and M. I. Bhangar, *Anal. Sci.*, **23(2)**, 193 (2007).
15. R. Soomro, M. J. Ahmed, N. Memon and H. Khan, *Analytical Chem. Insights*, **3**, 75 (2008).
16. H. Abdollahi, M. S. Panahi and M. R. Khoshayand, *Iranian J. Pharmaceut. Res.*, **207** (2003).
17. M. J. Ahmed and M. Nasir Uddin, *Chemosphere*, **67**, 2020 (2007).
18. N. Veerachalee, P. Taweema and A. Songsasen, *Kasetsart J.-Nat. Sci.*, **41(4)**, 675 (2007).
19. S. G. Prabhulkar and R. M. Patil, *Int. J. Chem. Sci.*, **6(3)**, 1480 (2008).

20. S. H. Guzar and Q. H. Jin, *Chemical Research in Chinese Universities*, **24(2)**, 143 (2008).
21. Y. Haoyi, Z. Guoxiu and Y. Gaohua, *Chem. Anal. Meterage*, 1 (2009).
22. B. Pal, B. Chowdhury and N. Triazene, *Mikrochimica, Acta*, **83**, 121 (1984).
23. D. D. Perrin and Boyd Dempsey, In *Buffers for pH and Metal ion Control*, Chapman and Hall Ltd., London (1974).
24. H. Eskandari, *Bulletin of the Korean Chemical Society*, **25(8)**, 1137 (2004).
25. A. Safavi and H. Abdollahi, *Microchem. J.*, **69**, 69 (2001).
26. K. Uchiyama, K. Ohsawa, Y. Yoshimura, T. W., K. Imaeda, *Analytical Sciences*, **8**, 655 (1992).
27. H. Abdollahi, M. S. Panahi and M. R. Khoshayand, *Iranian J. Pharmaceut. Res.*, 207 (2003).
28. H. B. Singh, N. K. Agnihotri and V. K. Singh, *Talanta*, **48**, 623 (1999).
29. H. H. Jaffe, Milton Orchin, *Theory and Applications of Ultraviolet Spectroscopy*, John Wiley and Sons Inc., New York (1962) p. 18.
30. M. A. Omary, O. Elbjeirami, C. S. P. Gamage, K. M. Sherman, H. V. R. Dias, *Inorganic Chemistry*, **48(5)**, 1784 (2009).
31. M. Thongngam and D. J. McClements, *Langmuir*, **21**, 79 (2005).
32. M. Vincekovic, Bujan. M. Smit. I. Filipovic V. N., *Colloids and Surfaces A*, **255**, 181 (2005).
33. Q. Gan, T. Wang, C. Cochrane and P. McCarron, *Colloids and Surfaces A*, **44**, 6 (2005).
34. M. I. Toral, N. Lara, J. Gomez and P. Richter, *Anal. Lett.*, **35(1)**, 135 (2002).
35. J. Yunn, Heeseon Choi, *Talanta*, **52**, 893 (2000).
36. G. A. Shar and G. A. Soomro, *The Nucleus*, **44(1-2)**, 33 (2007).

Revised : 20.05.2014

Accepted : 21.05.2014